

**SIRIM Berhad****INDUSTRIAL BIOTECHNOLOGY RESEARCH CENTRE, Building 19**

Tel: 03-55446953/6960

Fax: 03-55446988

**TEST REPORT**

REPORT NO: R207/13/B19/06

PAGE: 1 of 3

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

Applicant : Trumer Medicare Sdn Bhd.  
No. 16-2, Jalan SS19/1G,  
47500 Subang Jaya,  
Selangor.  
(Mr. KC Yap)

Manufacturer /  
Company : -

Sample : Nano Colloidal Argentum

Method of Test : Ocular Irritection<sup>®</sup> Test

Description of  
Sample : Received one sample for testing with the following identifications:

- a) Colour : Clear
- b) Shape/Form : Liquid
- c) pH : 5.5
- d) Quantity : Approximately 100 mL

Date Received : 21 June 2013

Job No. : J207/13

Issue Date : 26 JUL 2013

Approved signatories,

**(SUZAINI BADRUDIN)**

Researcher,  
Industrial Biotechnology Research Centre,  
SIRIM Berhad.

**(DR. NEELAM SHAHAB, AMIC)**

Senior Principal Researcher,  
Industrial Biotechnology Research Centre,  
SIRIM Berhad.

**SIRIM Berhad**

(No. Syarikat 367474 - V)  
1, Persiaran Dato' Menteri  
Seksyen 2, Peti Surat 7035  
40700 Shah Alam  
MALAYSIA  
Tel: 60-3-55446000  
Hotline: 60-3-55103535  
Faks: 60-3-55108095  
Website : [www.sirim.my](http://www.sirim.my)

## TEST REPORT

REPORT NO: R207/13/B19/06

PAGE: 2 of 3

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

### 1.0 Test timetable

Receipt of sample : 21 June 2013  
Study : 22 - 23 July 2013  
Report : 26 July 2013

### 2.0 Test method

#### 2.1 Study Objective

The Nano Colloidal Argentum provided by Trumer Medicare Sdn Bhd was evaluated with the Irritation<sup>®</sup> Assay System in order to predict its potential to cause ocular irritation.

To achieve this objective, standard volume-dependent dose-response study was performed using the Ocular Irritation<sup>®</sup> assay system.

#### 2.2 Background


The proprietary Ocular Irritation assay is a standardized and quantitative *in vitro* test which utilizes changes of relevant macromolecules to predict the acute ocular irritancy of chemicals and chemical formulations. This assay is based on the principle that chemical compounds will promote measurable changes in target biomolecules and macromolecular structures. Previous studies have clearly demonstrated that the processes of protein denaturation and disaggregation that are induced in this *in vitro* assay mimic the effects that are produced when these types of irritants are applied to the eye. Consequently, this *in vitro* test may be employed to predict the *in vivo* toxic effects of chemicals and formulations.

Additionally, the Ocular Irritation assay system provides significant benefits when compared to the *in vivo* Draize test method. The quantitative Ocular Irritation *in vitro* assay has been found to be highly reproducible. Of even greater relevance, the Ocular Irritation assay method can be readily employed to evaluate multiple samples at varying volumes or concentrations. Thus, the test serves as an extremely useful screening tool that facilitates all stages of raw material selection, formulation development and final product selection.

#### 2.3 Materials and Methods

The Ocular Irritation assay is a quantitative *in vitro* test method that mimics an acute ocular irritation test. To perform this standardized assay, the test sample is applied to a synthetic biobarrier composed of a semi-permeable membrane. Following application, the sample is absorbed by and permeates through this synthetic biobarrier to gradually come into contact with a proprietary solution containing glycoproteins. Reaction of the test sample with these proteins and macromolecular complexes promotes conformational changes that may be readily detected as an increase in the turbidity of the protein solution. The turbidity may be detected spectrophotometrically at a wavelength of 405 nm.

The irritancy potential of a test sample is expressed as an Irritation Draize Equivalent (IDE) score. This score is defined by comparing the increase in optical density (OD<sub>405</sub>) produced by the test material to a standard curve that is constructed by measuring the increase in OD<sub>405</sub> produced by a set of Calibration substances. These Calibrators have been selected for use in this test because their irritancy potential has been previously documented in a series of *in vivo* investigations. As a general guideline, the predicted ocular irritancy classification is shown in Table 1.

  
SUZAINI BINTI BADRUDIN  
Researcher  
Bioprocess Programme  
Industrial Biotechnology Research Centre  
SIRIM Berhad

## TEST REPORT

|   |              |
|---|--------------|
| REPORT NO: R207/13/B19/06   | PAGE: 3 of 3 |
| This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad. |              |

**Table 1.** Relationship of Irritation Draize Equivalent (IDE) Score to Irritancy Classification for the Ocular Irritation Test Method

| Irritation Draize Equivalent (IDE) | Predicted Ocular Irritancy Classification |
|------------------------------------|---|
| 0.0 - 12.5                         | Minimal Irritant                          |
| 12.5 - 30.0                        | Mild Irritant                             |
| 30.0 - 51.0                        | Moderate Irritant                         |
| 51.0 - 80.0                        | Severe Irritant                           |

All data were calculated and analyzed via a computer program which determines assay result acceptance based upon qualification parameters defined in the program. Irritancy score that correlates most closely with the *in vivo* irritancy properties of a test article is the highest qualified score calculated by the irritation software. This value is defined as the maximum qualified score.

A standard volume-dependent dose-response study was performed with the Ocular Irritation<sup>®</sup> test method.

### 3.0 Results and Discussion

Results and the predicted *in vivo* classifications analyzed via the Irritation software are shown in Table 2.


**Table 2.** Summary of Ocular Irritation<sup>®</sup> Results  
(<sup>a</sup> Maximum Qualified Score)

| Sample Description      | volume (μL) | IDE Score         | Predicted Ocular Irritancy Classification |
|-------------------------|-------------|-------------------|---|
| Nano Colloidal Argentum | 50          | 12.6              | Minimal/Mild Irritant                     |
|                         | 75          | 15.0              | Mild Irritant                             |
|                         | 100         | <sup>a</sup> 19.8 | Mild Irritant                             |

The Nano Colloidal Argentum was classified as minimal ocular irritant with an IDE score of 19.8.

### 4.0 Conclusion

The **Nano Colloidal Argentum** was evaluated for Ocular Irritation<sup>®</sup> and classified as **Mild Ocular Irritant**.

  
**SUZAINI BINTI BADRUDIN**  
Researcher  
Bioprocess Programme  
Industrial Biotechnology Research Centre  
SIRIM Berhad